

ScyTek
Laboratories

Instructions For Use
EAC-IFU

Rev. Date: Jan. 21, 2008

Revision: 1

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

EA-50 Stain Solution (Cytological Stain)

Description: EA-50 is a cytologic stain routinely used in diagnostic cytology to aid in the identification and classification of exfoliative cells. The stain is used for examining exfoliative cells of vaginal, cervical, sputum, as well as other body secretions. Commonly used in combination with hematoxylin.

Results: Nuclei: Blue
Cytoplasm: Various shades of Pink or Green.

Uses/Limitations: For In-Vitro Diagnostic use only.
Cytological applications.
Do not use past expiration date.
Use caution when handling these reagents.

Control Cells: Vaginal exfoliative cells
Cheek exfoliative cells

Storage: Room Temperature

Availability/Contents:

<u>Item #</u>	<u>Volume</u>
EAC500	500 ml
EAC999	1000 ml

Precautions: Flammable.
Keep away from open flame.
Avoid contact with skin and eyes.
Harmful if swallowed.
Follow all Federal, State, and local regulations regarding disposal.
Use in chemical fume hood whenever possible.

Storage: 18° C



25° C



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
435-755-9848
U.S.A.



Authorized Representative in Europe
(Regulatory affairs only)
EmergoEurope (31)(0) 70 345-8570
Molsnstraat 15
2513 BH Hague, The Netherlands

Instructions For Use EAC-IFU

Rev. Date: Jan. 21, 2008

Revision: 1

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

Procedure:

1. Fix exfoliated epithelial cell smears in ethanol.
2. Using graded alcohols, rehydrate to distilled water.
3. Stain slide in Hematoxylin (HMM) for 2-5 minutes.
4. Rinse for 2 minutes in several changes of tap water.
5. Dip slide in Bluing Reagent (BRT) for 30 seconds.
6. Rinse slide in distilled water.
7. Place slide in 95% alcohol for 2 minutes.
8. Stain slide in EA-50 Stain Solution for 2 minutes.
9. Rinse quickly in 95% alcohol.
10. Dehydrate in absolute alcohol.
11. Clear, and mount in synthetic resin.

References:

1. Liu, W. A simplified cytological staining technique. American Journal of Clinical Pathology, Volume 54, Page 767, 1970.
2. Keebler, C., Reagen, J. A Manual for Cytotechnology, American Society of Clinical Pathologists Press, Chicago, IL, Page 323, 1983.

Storage: 18° C



25° C



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
435-755-9848
U.S.A.



Authorized Representative in Europe



(Regulatory affairs only)
EmergoEurope (31)(0) 70 345-8570
Molsstraat 15
2513 BH Hague, The Netherlands