



Instructions For Use FAB-2-IFU		
Rev. Date: Aug. 17, 2014	Revision: 2	Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 – Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com


Acid Fast Bacteria (AFB) Stain Kit

Description: The Acid Fast Bacteria (AFB) Stain Kit is intended for use in the histological visualization of Acid Fast Bacteria and Tubercle Bacilli. This kit is a rapid 15-minute procedure. The lipid capsule of the acid-fast organism takes up carbol fuchsin and resists decolorization.

Acid Fast Organisms: Bright Red
 Background: Light Green

Uses/Limitations: Not to be taken internally.
 For In-Vitro Diagnostic Use.
 Histological applications.
 Do not use if reagents become cloudy.
 Do not use past expiration date.
 Use caution when handling reagents.
 Non-Sterile.

Control Tissue: Any well fixed paraffin embedded or frozen tissue section cut at 3-5µ.
 Cell Smear.




Ordering information regarding individual components on back page!

<u>Item #</u>	<u>Kit Contents</u>	<u>Volume</u>	<u>Storage</u>
CFZ030	Carbol Fuchsin Solution	30 ml	18-25°C
AAL030	Acid Alcohol Solution (0.5%)	2 x 30 ml	18-25°C
LGA030	Light Green Solution	30 ml	18-25°C

Precautions: Avoid contact with skin and eyes.
 Harmful if swallowed.
 Follow all Federal, State, and local regulations regarding disposal.
 Use in chemical fume hood whenever possible.
 Wear protective clothing.

Note: Carbol Fuchsin Solution should be filtered when a thick sheen develops on top of solution. To avoid possible contamination, gloves should be worn when performing this procedure. Do not use tap water prior to application of Carbol Fuchsin Solution as it is reported that Acid Fast Bacteria can be found in some systems. Therefore use of distilled water is recommended whenever possible.

Storage: 18° C  25° C

 ScyTek Laboratories, Inc.
 205 South 600 West
 Logan, UT 84321
 U.S.A.



 EmergoEurope (31)(0) 70 345-8570
 Molsstraat 15
 2513 BH Hague, The Netherlands

Procedure (Standard):

1. Deparaffinize sections if necessary and hydrate to distilled water.
2. Incubate slide in Carbol Fuchsin Solution for 15 minutes.
3. Rinse for 2 minutes in running tap water followed by 2 changes of distilled water.
4. Decolorize in Acid Alcohol Solution (0.5%) until sections are a pale pink color.
5. Dip slide twice in distilled water to rinse.
6. Counterstain in Light Green Solution for 1-2 minutes.
7. Rinse in several changes of distilled water.
8. Dehydrate quickly through graded alcohols ending with 2 changes in absolute alcohol.
9. Clear, and mount in synthetic resin.


References:

1. Sheenan, D.C., Hrapchak, B.B. Theory and Practice of Histotechnology, 2nd Edition. Battelle Press, Columbus, OH.
2. Carson, F.L., 1996, Histotechnology; A Self-Instructional Text, 2nd Edition. ASCP Press, Chicago, IL.

Bulk Reagent Ordering Information and Current Pricing at www.scytek.com

Description:	Catalog #	Volume
Carbol Fuchsin Solution	CFZ125	125 ml
	CFZ500	500 ml
	CFZ999	1000 ml
Acid Alcohol Solution (0.5%)	AAL500	500 ml
	AAL999	1000 ml
Light Green Solution	LGA125	125 ml
	LGA500	500 ml
	LGA999	1000 ml
	LGA3800	1 Gallon

Storage: 18° C  25° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.



 EmergoEurope (31)(0) 70 345-8570
Molsstraat 15
2513 BH Hague, The Netherlands