



Instructions For Use FMH-IFU

Rev. Date: Dec. 7, 2016

Revision: 2


Page 1 of 1

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 – Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

FluoreGuard Mounting Medium (Hard Set)

Description: FluoreGuard Mounting Medium (Hard Set) is a water-soluble medium designed for semi-permanent coverslipping of fluorescent slide preparations. FluoreGuard dries to a semi-rigid layer which eliminates tissue damage due to moving coverslips and facilitates long term storage. The (Hard Set) formulation provides better drying in humid conditions.

Uses/Limitations: Not to be taken internally.
For In-Vitro Diagnostic use.
Histological applications.
Do not use if reagents become cloudy.
Do not use past expiration date.
Use caution when handling reagents.
Non-Sterile.



Ordering Information and Current Pricing at www.scytek.com

Availability:	<u>Item #</u>	<u>Volume</u>
	FMH030	30ml Dropper
	FMH060	60ml Dropper
	FMH500	500ml
	FMH999	1000ml

Storage: Store at 18-25°C.

Precautions: Avoid contact with skin and eyes.
Harmful if swallowed.
Follow all Federal, State, and local regulations regarding disposal.


Procedure:

1. Apply several drops of FluoreGuard Mounting Medium (Hard Set) and coverslip.
2. Allow FluoreGuard to dry for 30 minutes prior to viewing. Product will dry to a semi-solid form overnight.
3. FluoreGuard is an aqueous based medium and may be removed by soaking in PBS overnight.


References:

1. Wanninger S, Donati M, Di Francesco A, Hässig M, Hoffmann K, Seth-Smith HM, Marti H, Borel N. Selective Pressure Promotes Tetracycline Resistance of Chlamydia Suis in Fattening Pigs. PloS one. 2016 Nov 28;11(11):e0166917.

Storage: 18° C  25° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.


Emergo Europe
Molenstraat 15
2513 BH Hague, The Netherlands